

The primary calibration was carried out by using the phase transitions observed by Bridgman⁽¹⁾ in bismuth at 24,700 atmospheres and in tellurium at 43,500 atmospheres. The secondary standard is the shift with pressure of the 2210 cm^{-1} vibration of CN^- ion dissolved by fusion in NaCl ⁽²⁾. This vibration shifts continuously to the "blue" (higher frequencies) with pressure. The data can be fitted by the equation

$$p = 0.965 \Delta\nu + 4 \times 10^{-3} (\Delta\nu)^2 + 5.3 \times 10^{-5} (\Delta\nu)^3$$

where $\Delta\nu$ refers to the change in frequency from the atmospheric pressure value in cm^{-1} and p is the pressure in thousands of atmospheres.

Cell I fails somewhere above 60,000 atmospheres because the moving carbide piston breaks in compression.

-
1. Bridgman, P. W., Proc. Amer. Acad. Arts and Sci., 74, 425, 1942.
 2. A detailed discussion of pressure effects on the CN^- stretching frequency will be published elsewhere.
-

Cell II

Cell II utilizes the same principles described for Cell I plus one further development. The pistons are much larger in diameter with a corresponding change in insert size, window holes, etc. Typical dimensions are summarized in Table I. The pistons are tapered with a flat section $3/32$ " in diameter in the center. (This is the smallest