

Technical Report Number 164

February 1968

Apparatus for Mechanical Testing of
Soft Crystals at High Pressure

S. H. Gelles

Abstract

Apparatus which has sufficient sensitivity for the tensile testing of soft single crystals at pressure levels to approximately 30 kbars is described. This equipment makes use of a tensile yoke and a load cell based on a linear variable differential transformer within the high pressure fluid. Strain is measured externally by monitoring the movement of the loading piston with a linear variable differential transformer and the load-displacement relationship is displayed on an X-Y recorder. Discussion of the calibration procedure and performance of the apparatus is included.

Ledgemont Laboratory

Kennecott Copper Corporation

Lexington, Massachusetts 02173