

LEDGEMONT LABORATORY
KENNECOTT COPPER CORPORATION
LEXINGTON, MASSACHUSETTS

Technical Number: TR-164 Dated: 26 Feb. 1968

Title Apparatus for Mechanical Testing
 of Soft Crystals at High Pressure
and

Author S. H. Gelles

Publisher: Submitted to:
 Rev. Sci. Instr. Pages: 30
 Illus. 13

Abstract: Apparatus which has sufficient sensitivity for the tensile testing of soft single crystals at pressure levels to approximately 30 kbars is described. This equipment makes use of a tensile yoke and a load cell based on a linear variable differential transformer within the high pressure fluid. Strain is measured externally by monitoring the movement of the loading piston with a linear variable differential transformer and the load-displacement relationship is displayed on an X-Y recorder. Discussion of the calibration procedure and performance of the apparatus is included.