

34. Pipe fittings (galvanized) as follow:

- a. street elbows, 3/8 in., 6 ea.
- b. elbows, 3/8 in., 4 ea.
- c. short nipples, 3/8 in., 4 ea.
- d. Tees, 3/8 in., 2 ea.
- e. cross, 3/8 in., 1 ea.
- f. elbows, 1/2 in., 4 ea.
- g. street elbows, 1/2 in., 2 ea.
- h. tees, 1/2 in., 2 ea.
- i. short nipples, 1/2 in., 4 ea.
- j. 3/8 in. pipe, galvanized, 1 length.
- k. 1/2 in. pipe galvanized, 1 length.

35. Parker, Push-lok-82 fittings and hose as follows:

- a. Push-lok 3/8 - 3/8 - 30182-4, 8 ea.
- b. Push-lok 3/8 - 6 - 30682-4, 8 ea.
- c. Push-lok hose # 3/8-831, 20 ft.
- d. Push-lok 1/2-1/2-30182-4, 6 ea.
- e. Push-lok 1/2 - 8 - 30682-4, 6 ea.
- f. Push-lok hose # 1/2 -831, 20 ft.

36. Tubing to pipe adapter, 3/8 NPT to 3/8 -37°, straight through,
Parker no. 6-6-0103-1, 8 ea.

37. Tubing to pipe adapter, 1/2 NPT to 1/2 - 37°, straight through,
Parker no. 8-8-0103-1, 8 ea.

38. Gauge: Heise, 12 in., 12,000 psi max., with automatic thermal
compensator and slotted link, port opening in back of case to connect
to Autoclave Engineers 5/16 - 9/16 in. "Slimline" tubing for 15,000 psi
use. Gage divisions 0-12,000 in steps of 10 psi.